

Ambient Air Monitoring (using diffusion tubes)

Parameter monitored	Standard methodology used	Detection limit (as mass on diffusion tubes)	Is this company accredited for this test? ¹
Nitrogen dioxide	Colorimetric using UV/VIS spectrophotometry	0.065 µg	Yes
Sulphur dioxide	Ion Chromatography	0.04 µg	Yes
Ozone	Ion Chromatography	0.088 µg	Yes
BTEX	Gas Chromatography / FID	1.9 ng	Yes

¹ Accredited tests are subcontracted