

## Microbiological contamination

Parameter monitored	Standard methodology used	Medium	Detection limit	Is this company accredited for this test? <sup>1</sup>
Yeasts and moulds	ISO 18593:2004, ISO 21527-1:2008, ISO 18593:2004, ISO 21527-2:2008	Indoor and outdoor air, surfaces, waters.	1 cfu	Yes
Colony counts at 30°C	ISO 6222:1999	“	1 cfu	Yes
Colony count at 22°C	ISO 6222:1999	“	1 cfu	Yes
Escherichia coli	ISO 9308-1:2000	“	1 cfu	Yes
Staphylococcus aureus	ISO 18593:2004	“	1 cfu	Yes
Total coliforms	ISO 9308-1:2000	“	1 cfu	Yes
Faecal coliforms	ISO 9308-1:2000	“	1 cfu	Yes
Enterococchi	ISO 7899-2:2000	“	1 cfu	Yes
<i>Clostridium perfringens</i> (spores)	DLgs N° 31 02/02/2001 All III	“	1 cfu	Yes

<sup>1</sup> Accredited tests are subcontracted