

Analysis of particulate matter filters

Parameter monitored	Standard methodology used	Detection limit	Units	Is this company accredited for this test?
Al	OENORM EN ISO 17294	26	ng/m ³	No
As	OENORM EN ISO 17294	0.3	ng/m ³	No
Cd	OENORM EN ISO 17294	0.05	ng/m ³	No
Cr	OENORM EN ISO 17294	2	ng/m ³	No
Cu	OENORM EN ISO 17294	5	ng/m ³	No
Fe	OENORM EN ISO 17294	11	ng/m ³	No
Pb	OENORM EN ISO 17294	1	ng/m ³	No
Mn	OENORM EN ISO 17294	0.8	ng/m ³	No
Ni	OENORM EN ISO 17294	0.6	ng/m ³	No
Sn	OENORM EN ISO 17294	0.6	ng/m ³	No
V	OENORM EN ISO 17294	0.5	ng/m ³	No
Ammonium	ISO 14911	0.004	µg/Nm ³	No
Calcium	ISO 14911	0.003	µg/Nm ³	No
Magnesium	ISO 14911	0.003	µg/Nm ³	No
Potassium	ISO 14911	0.003	µg/Nm ³	No
Sodium	ISO 14911	0.002	µg/Nm ³	No
Chloride	EN ISO 10304	0.003	µg/Nm ³	No
Nitrate	EN ISO 10304	0.002	µg/Nm ³	No
Sulphate	EN ISO 10304	0.003	µg/Nm ³	No